

PRINTED: 09/06/2023
FORM APPROVED
OMB NO. 0938-0391

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE	TITLE	(X6) DATE
---	-------	-----------

FORM CMS-2567(02-99) Previous Versions Obsolete Event ID: 5RY111 Facility ID: 1906 If continuation sheet Page 1 of 3

DEPARTMENT OF HEALTH AND HUMAN SERVICES
CENTERS FOR MEDICARE & MEDICAID SERVICES

PRINTED: 09/06/2023
FORM APPROVED
OMB NO. 0938-0391

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 205115	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____		(X3) DATE SURVEY COMPLETED C 08/30/2023
NAME OF PROVIDER OR SUPPLIER DEXTER HEALTH CARE			STREET ADDRESS, CITY, STATE, ZIP CODE 64 PARK STREET DEXTER, ME 04930		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)		(X5) COMPLETION DATE
F 689	Continued From page 1 size sling and failed to properly apply the Hoyer lift sling resulting in R1 sliding out of the sling to the floor. R1 landed on the Hoyer lift legs causing a 9 centimeter (cm) long by 1.7 cm wide abrasion on his/her back. R1 was sent to the hospital for evaluation and returned to facility with no injuries other than the abrasion. On 8/30/23, R1's clinical record was reviewed. A discharge report from the hospital indicated the resident had no fractures and was discharged with a diagnosis of superficial abrasion on the back. Nurse's notes dated from 8/20/23 thru 8/22/23 indicated that the R1 did not complaint of pain in the area of the abrasion. On 8/30/23, the facility's "Lifting Machine-Using a Portable" policy and procedure was reviewed. The policy indicated under "Steps in the Procedure" #3: Select a sling that is appropriate for the resident's size and the task. #10: Place the sling under the resident. Visually check the size the ensure it is not too large or too small. #12: Attach sling straps to sling bar, according to manufacturer's instructions. On 8/30/23 at 8:55 a.m., interview with the Director of Nursing (DON). DON stated an investigation was started immediately after the incident occurred. In her interview with CNA1, she stated CNA1 told her that she used the wrong size Hoyer sling and CNA1 stated she did not crisscross the straps of the sling around R1's legs. The DON also stated that CNA2 told her that the sling was the wrong size and CNA1 did not cross the straps. The DON stated that for this type of sling, the sling straps must be crisscrossed around the resident's legs to prevent slipping out	F 689			

DEPARTMENT OF HEALTH AND HUMAN SERVICES
CENTERS FOR MEDICARE & MEDICAID SERVICES

PRINTED: 09/06/2023
FORM APPROVED
OMB NO. 0938-0391

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 205115	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____		(X3) DATE SURVEY COMPLETED C 08/30/2023
NAME OF PROVIDER OR SUPPLIER DEXTER HEALTH CARE			STREET ADDRESS, CITY, STATE, ZIP CODE 64 PARK STREET DEXTER, ME 04930		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETION DATE	
F 689	Continued From page 2 of the sling. The DON stated a program was promptly put in place to re-train all staff on the proper use of the Hoyer lift, sling size and sling application. The training was completed on 8/25/23. On 8/30/23 at 10:44 a.m., in an interview with the surveyor, CNA1 confirmed she used the wrong sized sling and that she assumed it was the correct size, and she did not cross the straps. On 8/31/23 at 1:45 p.m., in an interview with the surveyor, CNA2 stated that she was trained that the sling needed to be crisscrossed around the resident's legs. CNA2 confirmed that CNA1 did not crisscross the straps.	F 689			