

PRINTED: 12/27/2016
FORM APPROVED

Healthcare Licensing and Surveys

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: WY0028	(X2) MULTIPLE CONSTRUCTION A. BUILDING: _____ B. WING: _____	(X3) DATE SURVEY COMPLETED R 12/08/2016
NAME OF PROVIDER OR SUPPLIER SHERIDAN MANOR		STREET ADDRESS, CITY, STATE, ZIP CODE 1861 BIG HORN AVE SHERIDAN, WY 82801		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETE DATE
(S7999)	State Miscellaneous Life Safety State Miscellaneous This STANDARD is not met as evidenced by: Based on observation and staff interview, the facility failed to provide exhaust requirements in accordance with Chapter 3 Wyoming Rules and Regulations for Health Care Facilities. The findings were: Observation on 12-6-2016 at 2:25 PM adjacent to room 303 revealed that the exhaust fan was to be repaired or replaced by their previous compliance date of 11-02-2016. Further observation revealed that when inspected the fan was not running. Interview with Facility Staff indicated that the fans were replaced and were unaware why the exhaust fan was not in working order. Ref: Chapter 3 Wyoming Rules and Regulations for Health Care Facilities Section 5 b (iv) (B)	(S7999)	S7999 State Miscellaneous Life The facility is committed to providing working ventilation. Corrective Action: The exhaust fan on the 300 hall was replaced on 09/22/16 and when the state came in for follow-up visit, it was found to not be working. A repairman was called in and upon investigating; it was found that it was faulty pulley and broken belt. The pulley and belt were replaced and are now in working order. ID of Others: An audit of all exhaust fans were completed by the NHA and Maintenance Director. All are in working order. Systemic Changes: The NHA/Designee will educate the Maintenance Director on routine checking and maintenance of the exhaust fans.	1/17/17 1/17/17 <i>[Signature]</i>

Wyoming Dept of Health, Aging Division, Healthcare Licensing and Surveys
LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE

TITLE

(X6) DATE

STATE FORM

6400

CHVM12

If continuation sheet 1 of 1

RECEIVED

JAN 10 2017

Healthcare Licensing and Surveys

POC ACCEPTED VIA EMAIL TO DEREK HOLIBROOK ON
1-18-17.

PRINTED: 12/27/2016
FORM APPROVED

Healthcare Licensing and Surveys

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: WY0028	(X2) MULTIPLE CONSTRUCTION A. BUILDING: _____ B. WING: _____	(X3) DATE SURVEY COMPLETED R 12/06/2016
NAME OF PROVIDER OR SUPPLIER SHERIDAN MANOR		STREET ADDRESS, CITY, STATE, ZIP CODE 1851 BIG HORN AVE SHERIDAN, WY 82801		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETE DATE
{S7999}	State Miscellaneous Life Safety State Miscellaneous This STANDARD is not met as evidenced by: Based on observation and staff interview, the facility failed to provide exhaust requirements in accordance with Chapter 3 Wyoming Rules and Regulations for Health Care Facilities. The findings were: Observation on 12-8-2016 at 2:25 PM adjacent to room 303 revealed that the exhaust fan was to be repaired or replaced by their previous compliance date of 11-02-2016. Further observation revealed that when inspected the fan was not running. Interview with Facility Staff indicated that the fans were replaced and were unaware why the exhaust fan was not in working order. Ref: Chapter 3 Wyoming Rules and Regulations for Health Care Facilities Section 6 b (iv) (B)	{S7999}	Monitoring: The Maintenance Director/Designee will complete audits for 3 months making sure exhaust fans are in working order.	1/3/17

Wyoming Dept of Health, Aging Division, Healthcare Licensing and Surveys
LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE

TITLE

(X6) DATE

STATE FORM

CHVM12

If continuation sheet 1 of 1

RECEIVED

JAN 10 2017

Healthcare Licensing and Surveys